

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 27 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130288</u> ✓	
A	Name <u>DIXIE SPENCER</u>	ID# <u>2280</u>	Date <u>5/18/09</u> ✓
	Agency <u>AST ANIAK</u>	Phone # <u>675-4398</u>	
	Instrument Location <u>ANIAK</u> ✓		
B	Alco S/N <u>56799</u> ✓ Target Value <u>.084</u> ✓ High Pressure <u>1100</u> ✓		
	Alco Test Values 1 st Alco <u>.085</u> ✓ 2 nd Alco <u>.084</u> ✓		
	Signature <u>Dixie Spencer</u> ✓		
(OVER)			

BMB
5/28/09

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 1st day of June, 2009.

Carolyn M. Noland
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130288 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130288

MAY 18, 2009

OPERATOR'S NAME:

DIXIE W SPENDER

OPERATOR'S NUMBER: 2280

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI:

TEST

O.L. #: 123456

DEPT/AGENCY: AN10

CASE/REPORT: 09-00000

TEST TYPE: U

ALCO TARGET VALUE: .004 ✓

ALCO S/N: 56789

--- BREATH ANALYSIS ---

.004 ADJUSTED FOR 29.86 in ✓

ALCO TARGET .003 12:43

BLANK TEST .000 12:50

INTERNAL STANDARD .005 12:50

ALCO TV 29.86 in .005 12:51

BLANK TEST .000 12:51

SUBJECT SAMPLE .000 12:52

BLANK TEST .000 12:53

ALCO TV 29.86 in .004 ✓ 12:53

BLANK TEST .000 12:54

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130288

MAY 18, 2009

TIME 12:55

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

WARRANTY DATE: 02/20/01

FILE CHANGER: 40c

DATA TUBE: 41c

BAROMETER: 29.88 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!##%&'()*+,-./0123456789:;<=>?@AB CDEFG
HIJKLMN O PQRSTU VWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~